

CT DESPP Division of Scientific Services

Draeger 9510 In-House Certification Test Results

Instrument Serial Number: _____

Analyst: _____ Date: _____

Materials/Reagents:

0.02 g% Dry Gas

Lot Number: _____ Expiration Date: _____

0.08 g% Dry Gas

Lot Number: _____ Expiration Date: _____

0.25 g% Dry Gas

Lot Number: _____ Expiration Date: _____

Certified Alcohol Reference Solution for Simulator

Lot Number: _____ Expiration Date: _____

Methanol Solution for Simulator

Lot Number: _____ Expiration Date: _____

Results:

Test	Result 1	Result 2	Acceptable Result
0.02 g% Verification			0.0150-0.0250 g%
0.08 g% Verification			0.0750-0.0850 g%
0.25 g% Verification			0.2375-0.2625 g%
Blank 0.000		N/A	0.000 g%
Ethanol Detection	☐ Y ☐ N	N/A	Ethanol detected
Methanol Interferent	☐ Y ☐ N	N/A	Interferent detected

In-House Certification Final Result:

☐ PASS

☐ FAIL

Comments: