Balance ID:	
Mass Reference Standard Serial Number:	

	Actual weight within limits (+/- 1% of expected)				
Date	Check Weight Expected Value	Check Weight Actual Value	Yes	No*	Initials
2000	22250000 10000	1100001			
			IV		

Balance is checked each day of use with one calibrated weight in the approximate range of the samples being weighed.

<sup>\*</sup> If the check weight is not within acceptable limits see the section supervisor.