

*Approved by Director: Dr. Guy Vallaro***A. Purpose:**

The focus of the National Integrated Ballistic Information Network (NIBIN) is to provide federal, state and local law enforcement agencies with investigative leads by comparing digital images of fired cartridge cases from different crime scenes. The NIBIN database utilizes BrassTrax HD3D, an automated acquisition station for capturing images of fired cartridge cases, and MatchPoint+, a viewing station for reviewing correlation results. High-resolution images captured by the BrassTrax HD3D acquisition station are stored on regional servers maintained by the Bureau of Alcohol, Tobacco and Firearms and Explosives (ATF). Utilizing MatchPoint+, an analyst can review the correlation results and compare the images side by side on a computer monitor. If a potential association (a hit) is located during the review process, the associated agency(ies) are notified via a NIBIN Quick Hit (NQH) memo. If the agency desires, confirmation of the hit can be conducted by an examiner using a comparison microscope.

ATF Minimum Required Operating Standards (MROS)**1. Quality Assurance Program**

DSS has a NIBIN terminal in its facility and DSS has a documented quality system that is appropriate to the NIBIN acquisition and correlation processes that comply with ASCLD-LAB and ISO/IEC 17025 and the ATF Standards. The quality system will include the following:

- Goals and Objectives as in General Laboratory SOPs along with the Firearms Unit SOPs (Standard 1.1.1.1)
- Organization and Management as represented in GL-1 (Standard 1.1.1.2)
- Personnel as represented in the General Laboratory SOPs which include training, ethics, safety and security (Standard 1.1.1.3)
- Acquisition, Correlation, NIBIN Lead Dissemination Procedures and articulated in Section B.1.g of this SOP (Standard 1.1.1.4)
- Review of casework prior to release of results as designated in GL18 and FA-SOP-06 (Standard 1.1.1.5)
- Corrective Actions and Audits as described in GL-9 (Standard 1.1.1.6)

2. Organization and Personnel

NIBIN Program Administrator: full-time examiner employed by the Division of Scientific Services the NIBIN site. This individual is designated to communicate with all parties (i.e submitting law enforcement agencies, ATF Crime Gun Intelligence Centers {CGICs}, ATF

Approved by Director: Dr. Guy Vallaro

NIBIN Unit) involved in the NIBIN process. The NIBIN Program Administrator will work with the Case Management Unit and Deputy Director in communicating to the submitting agencies. The NIBIN Program Administrator must be a qualified NIBIN user in good standing. The NIBIN Program Administrator will work in conjunction with Top Management and the Quality Section in implementing and directing policies, procedures and operations of the NIBIN site along with any subsequent audits. (Standard 2.1)

In the event that the NIBIN Program Administrator position is vacated and there is no employee present who meets the requirements of the above standard, DSS will immediately contact the ATF and submit a contingency plan to the ATF within 14 days for approval. No new casework involving the use of NIBIN can be started until such plan is approved. (Standard 2.1.2.4.1)

If the NIBIN Program Administrator position is vacated, the Deputy Director will notify the Director and Quality Manager. The Director and Deputy director will, within 7 days of the NIBIN Program Administrator's position being vacated, designate an individual from the Firearm Unit as the Interim NIBIN Program Administrator. A memo from DSS Laboratory Management will be generated informing the appointee of the appointment. The Interim NIBIN Program Administrator must meet the requirements as listed in Standard 2.1.

The ATF will be contacted with the name and qualifications of the named Interim NIBIN Program Administrator. Once the contingency plan is approved, this individual will named the NIBIN Program Administrator.

NIBIN Users: Assigned Unit personnel or other laboratory-authorized individuals who have successfully completed an ATF-approved acquisition and correlation course provided by the ATF, Forensic Technology, Inc., and/or a NIBIN Authorized Trainer to use the equipment to perform Acquisition and/or Correlation. (Standard 2.1.3.1)

Records of training, qualifications, skills and experience of the NIBIN Administrator and NIBIN Users are maintained by the DSS Quality Management System (Standard 2.1.4 and GL 11)

3. Facility and Security of NIBIN Equipment:

DSS will comply with all ATF, DOJ and/or Federal security requirements related to the NIBIN program including its network and systems to ensure the integrity of the NIBIN analysis as well as the evidence. (Standards 3.1 / 3.1.1)

DSS will contain the NIBIN equipment in a room that has a locked door and access limited to personnel assigned to the Firearms Unit or other employees as designated by the Director. DSS has proximity cards and keys assigned to all employees. The proximity cards have permission levels that are set that limit access to the NIBIN area as designated by the NIBIN Administrator and Management. (Standards 3.1.2 / 3.1.6 / 3.1.7 / GL-3)

Approved by Director: Dr. Guy Vallaro

Employees no longer authorized to use the NIBIN equipment will be limited to room containing the NIBIN equipment via a change in the proximity access level and/or removal of the keys. (Standard 3.1.8 / GL-3)

DSS is a secure facility that has security monitoring and limited access via proximity cards. This facility is alarmed when employees are not in the building. Main doors to the operational areas are locked and require an appropriate proximately card to acquire access. (Standards 3.1.3 / 3.1.4 / 3.1.6 / GL-3)

Visitors to DSS are required to follow the guidance as provided in GL 3 “Security” Section 2 “DSS Facility Access to Visitors”. (Standard 3.1.9)

DSS accepts NIBIN related evidence through its Evidence Receiving Unit. Submitting agencies must buzz into the Evidence Receiving area to enter the location. This procedure is outlined in GL-3 Section 1. B. ii. (Standard 3.1.5)

4. Evidence Control:

The Firearms Unit uses GL-13 “General Evidence Handling” as guidance on the proper handling of evidence to main the forensic integrity of case materials. Evidence in this unit will be handled in a manner that appropriately documents custody and storage locations. (Standard 4.1 / 4.1.3)

Evidence submitted to DSS is given a unique identifier as to the case number submitted to the case along with a unique identifier for the evidence submitted. Guidance for the case numbering, submission numbering and evidence tracking is clarified in GL-4 “LIMS”. The evidence when submitted to DSS is given a barcode that is used for tracking of this evidence to personnel or evidence storage locations. (Standard 4.1.1)

The Chain of Custody for the items of evidence submitted for NIBIN entry are tracked electronically via the barcode of the evidence and a unique PIN for each examiner. A record of the chain of custody is maintained in the Justice Trax database and can be printed when needed. (Standard 4.1.2)

Access to evidence locations are limited to personnel authorized by the Director or his designee. This access is designated in the Justice Trax permissions and only authorized personnel can transfer the evidence into or out of specific storage locations. (Standard 4.1.4)

B. Procedure:

1. NIBIN Entry Protocols

a. Exhibits that may be entered:

- i. Cartridge cases fired in semiautomatic and fully automatic pistols.

Approved by Director: Dr. Guy Vallaro

- ii. Cartridge cases fired in semiautomatic, fully automatic, bolt action, pump/slide action and lever action rifles.
- iii. Shotgun shells fired in semiautomatic, slide/pump action, lever action and bolt action shotguns.
- iv. Any firearm that is magazine fed and/or designed to eject cartridge cases after firing.

b. Exhibits that should not be entered:

- i. Cartridge cases or shotgun shells that lack sufficient or repeatable characteristics within the breech face and firing pin impressions to assist in the correlation review. The analyst's discretion will be used to make this determination.
- ii. Images of cartridge cases or shotgun shells fired in revolvers, break open shotguns, derringers and single shot firearms.
- iii. Evidence cartridge cases that are consistent with having been fired in the above type weapons.
- iv. Any law enforcement officer's duty weapon, unless extenuating circumstances apply.

2. Entering Case Information into BrassTrax HD3D**a. Creating case records**

- i. The procedure outlined in the BrassTrax HD3D users guide will be utilized to create a new case record and to enter the case information.
- ii. Case ID – The “DSS” or “ID” number section or the eleven digit Weapon Destruction Unit number. A dash (-) separating DSS or ID and the case number digits is optional.
Examples: DSS13-1234, ID-13-9876, or 40001652345
- iii. Event Type – Will be selected based on the information provided by the submitting agency. The event type should be the offense associated with the case.
- iv. Occurrence Date – The date of offense provided by the submitting agency.
- v. Case Supervisor – The analyst responsible for the case.
- vi. Originating Agency – The police agency submitting the evidence. If the agency is not listed, a note of the submitting agency should be made in the comment field.
- vii. Originating Agency Contact – The name of the lead investigator listed in the case.
- viii. Originating Agency Reference – The submitting police agency's case number.
- ix. High Profile – Shall be used at the discretion of the analyst.

- x. Comment – Any additional comments about the case can be entered at the discretion of the analyst.

b. Creating cartridge case exhibit records

- i. The procedure outlined in the BrassTrax HD3D users guide will be utilized to create a new cartridge case exhibit and to enter the exhibit information.
- ii. Exhibit number:
- Cartridge cases from an unknown firearm will use the case submission number or the number designated by the analyst.
 - Test fired cartridge cases will include the submission number of the firearm, the test fire number, and “TFCC”.

- iii. The following fields will be added in the cartridge case exhibit screen:

Caliber

Make of ammunition

Firing Pin shape

Firearm exhibit number (if applicable)

Composition (composition of cartridge case: nickel-nickel, brass-nickel, etc.)

Breech face characteristics

Category:

- Crime Evidence (items collected from a crime scene)
- Test Fire – Returned (use for firearms which will be returned to their rightful owner, or otherwise released from law enforcement custody)
- Test Fire – Terminated (use for all other firearms). If test fires are obtained from a firearm submitted in the case, the event type for the cartridge case exhibit will be changed to “test fire”.

c. Creating firearm exhibit records

- i. The procedure outlined in the BrassTrax HD3D users guide will be utilized to create a new firearm exhibit and to enter the exhibit information.
- ii. Firearm exhibit number - Use the submission number of the weapon or the number designated by the analyst.
- iii. The following fields will be added in the firearm exhibit screen:

Caliber

Approved by Director: Dr. Guy Vallaro

Make

Model (if known)

Type

Serial Number

- If obliterated and later restored, the serial number field shall be updated.
- If the serial number could not be restored, enter “Obliterated” or “Oblit.”
- If the firearm does not have a serial number, enter “N/A” or “none.”

d. Acquisition Parameters

i. Centerfire Cartridge Cases/Shotshells

- a) The analyst will acquire the following regions of interest:

Breech Face

Full Headstamp

Firing Pin

Ejector marks (if present)

ii. Rimfire Cartridge Cases

- a) The analyst will acquire the following regions of interest:

Full Headstamp

Firing Pin

iii. Other Acquisition Parameters

- a) The analyst will follow the BrassTrax HD3D acquisition protocols outlined in the BrassTrax HD3D users guide to orient the cartridge case/shotgun shell and capture the digital images.
- b) If a cartridge case/shotgun shell is oriented outside of the above parameters, a remark shall be made in the comment field stating the reason for the deviation.
- c) In the event that reference collection parts (e.g. firing pin, slide, etc.) are used to obtain test fired cartridge cases/shotgun shells, the images of the respective areas need not be captured. A remark shall be made in the comment section describing which parts were replaced.

e. Correlation Results

- i. Acquired exhibits will automatically correlate against the statewide database. In the event an agency requests a search of other geographical locations, the analyst will perform a manual correlation for the region requested.
- ii. Using the Rank Sort feature, the analyst will review at minimum the first fifty (50) images of the Top 20 best scored results of the correlation.
- iii. Once the correlation results have been reviewed and there are no hits, the Top 20 Rank Sort results will be printed and kept in the case file.
- iv. The correlation request may be deleted from the correlation results list.

f. Generating Hits

- i. If the analyst reviewing the correlation results locates a potential hit, they will mark the cases as an unconfirmed hit according to the MatchPoint+ users guide protocols after a second analyst has reviewed the potential hit.
- ii. If a hit has been microscopically confirmed by a firearms examiner, the analyst will mark the hit as “confirmed,” using the steps outlined in the MatchPoint+ users guide. The analyst will note in the comments field who confirmed the hit.

g. Case Records and Reports

- i. A preliminary NIBIN Quick Hit memo (NQH) will be issued for unconfirmed hits.
 - a) The NQH memo will be technically and administratively reviewed by the Unit supervisor or designee prior to being released.
 - b) The original NQH will be furnished to the main case file located in the Evidence Receiving Unit.
 - c) A copy of the NQH will be sent to the submitting agencies involved with the unconfirmed hit.
 - d) A copy of the NQH will be maintained in the firearms case file.
 - e) A notation in JusticeTrax will be made, stating the case numbers involved and the date the NQH memo was sent.
 - f) If the unconfirmed hit involves a destruction gun from the Special Licensing and Firearms Unit, the analyst will fill out the Weapon Destruction Unit form letter and attach it to the NQH memo.
 - g) If a candidate for comparison is not found, the case can remain in the custody of the laboratory until either the case is completed, or a report is issued stating that NIBIN entry was made, no hits were developed, the evidence is being returned to the

Approved by Director: Dr. Guy Vallaro

submitting agency unworked, and if further examination is required, to contact the laboratory.

h. NIBIN Hit Maps

- i. If a hit is related to four (4) or more cases, a NIBIN hit map will be created using a word processing program and will be saved in the folder “NIBIN Hit Maps” located on the network drive.
- ii. The map will be part of the NIBIN Quick Hit memo and is subject to review.
- iii. Each box on the map shall contain the laboratory case number, the agency and agency case number, the type of crime, and the caliber(s) of submitted cartridge cases.
 - a) Firearms submitted in the case will have the caliber/make/model/serial number listed.
 - b) Different colored fonts and arrows will represent cartridge cases that share/may share a common origin.
 - c) Dotted arrows will indicate unconfirmed hits. Solid arrows will indicate confirmed hits.
 - d) Maps will be updated as additional hits are made and/or hits are confirmed on the microscope.
 - e) Updated maps will be added to the associated case jackets as additional documentation.

i. Maintenance and Quality Assurance

- i. The NIBIN monthly check will be completed, by the fifth of each month, by the assigned analyst. The Top 20 Rank Sort correlation results and case information pages will be printed and stored in the NIBIN Monthly Control Log binder located in the NIBIN room. Refer to FA SOP-33 Equipment Maintenance Schedule for instructions.
- ii. Tests and Calibrations activities are maintained in the BrassTrax HD3D software. This report can be generated at will.

C. Equipment:

1. BrassTrax HD3D acquisition station
2. MatchPoint+ viewing station

D. References:

1. BrassTrax HD3D users guide (located on the BrassTrax computer)

**State of Connecticut Department of Emergency Services and Public Protection
Division of Scientific Services**

Documents outside of Qualtrax are considered uncontrolled.

Approved by Director: Dr. Guy Vallaro

2. MatchPoint+ users guide (located on the MatchPoint+ computer)
3. FA SOP-33 Equipment Maintenance Schedule

ARCHIVED