

**FA SOP-33 Equipment Maintenance Schedule**

Document ID: 1140

Revision: 6

Effective Date: 6/4/2018

Status: Published

Page 1 of 2

*Approved by Director: Dr. Guy Vallaro*

All instruments and equipment in the Firearms Unit shall be calibrated and maintained on a set schedule. Calibration and/or cleaning of equipment shall be conducted by a laboratory-approved vendor.

Maintenance and calibration records shall be maintained on all instruments and equipment. Records shall be kept in a log in the appropriate area of the Firearms Unit.

<b>Equipment Type</b>	<b>Schedule</b>
Comparison Microscopes	1, 8, 9
Stereo Microscopes	1
Balances	1, 2
Barrel/Overall Measuring Device	2, 4
Calipers	2, 4
NRA Static Weights	4, 5
Spring Gage	2, 3
NIBIN Acquisition Unit	2, 9
Metric Weights Set	4, 5
Gauge Blocks	4
Indoor Range	6
Vertical Water Tank	7
Stage Micrometers	4

*\* The NRA static weights and the metric weights set are traceable to National Institute of Standards and Technology (NIST) standards.*

1. Cleaned annually by lab-approved vendor
2. Checked monthly and recorded in appropriate log book; to be done by 5<sup>th</sup> of every month
3. Out for calibration every 3 years by lab-approved vendor; sooner if damage occurs
4. Out for calibration every 5 years by lab-approved vendor; sooner if damage occurs
5. Checked annually in-house
6. Cleaned every 3 years by lab-approved vendor
7. Water changed monthly or as needed depending on use
8. Magnification checked bi-annually with stage micrometers
9. Run equipment-specific QC check when vendor upgrades software

*Approved by Director: Dr. Guy Vallaro*

**References:**

1. FA SOP-35 NIBIN Quality Control Check
2. FA SOP-39 Microscope Quality Control Check
3. Firearms Unit Uncertainty of Measurement Budget
4. FA QR-16 Firearms Unit Equipment List

ARCHIVED